



National Institute of Standards & Technology

Certificate of Analysis

Standard Reference Material 3130a

Spectrometric Standard Solution

Lutetium

Batch Code 391712

This Standard Reference Material (SRM) is intended for use in atomic absorption spectrometry, optical emission (plasma) spectrometry, spectrophotometry, or any other analytical technique that requires aqueous standard solutions for calibrating instruments. SRM 3130a is a single element solution prepared gravimetrically to contain a nominal 10 mg/mL of lutetium with a nitric acid concentration (V/V) of 10 percent. The density of the solution at 22 °C is 1.115 g/mL.

Metal	Concentration (mg/mL)	Source Purity, %	Acid Conc. (V/V) Approximate
Lu	9.98 ± 0.04	Lu_2O_3 (99.99) ^a	HNO_3 , 10%

^aThis high-purity material was analyzed by atomic absorption spectrometry and optical emission spectrometry and found to contain less than 100 µg/g total impurities.

The certified value (V) is based on replicate titrations against a reference solution of Lu metal of known purity. The value has been adjusted upward by 0.1% relative, based on estimated transpiration losses of solvent through the container walls of 0.2% relative per year after the bottle is removed from the plastic sleeve.

The uncertainty in the certified value is calculated as:

$$U = (2u_c + 0.001V) \text{ mg/mL}$$

where u_c is the "combined uncertainty" calculated according to the CIPM approach. [1] The value of u_c is intended to represent, at the level of one standard deviation, the combined effect of uncertainty components associated with volumetric, gravimetric, and titrimetric factors, as well as the purity of the Lu metal. The additional quantity, 0.001V, is an allowance for transpiration of the solution through the container walls, which is estimated to be $\pm 0.1\%$ of the certified value during the one-year period of validity of the certification.

SRM 3130a was prepared by T.A. Butler; atomic absorption and emission spectrometric analyses and titrimetric analyses were made by T.A. Butler, C.M. Beck II, and J.A. Norris of the NIST Inorganic Analytical Research Division.

The technical and support aspects involved in the preparation, certification, and issuance of this SRM were coordinated through the Standard Reference Materials Program by J.S. Kane.

Gaithersburg, MD 20899
January 11, 1994

Thomas E. Gills, Acting Chief
Standard Reference Materials Program

(over)

Procedures for Use

Stability: This certification is valid for one year from the shipping date, provided the solution is kept tightly capped and stored under normal laboratory conditions. NIST will monitor the stability of representative solutions from the SRM lot; and if any changes occur that invalidate this certification, NIST will notify purchasers.

Preparation of Working Standard Solutions: All solutions should be brought to 22 ± 1 °C before use and all glass or plastic surfaces coming into contact with the standard must have been previously cleaned. A working standard solution can be prepared from the SRM solution by serial dilution. Dilutions should be made with certified volumetric class A flasks and 5 or 10 mL class A pipets. All volumetric transfers of solutions should be performed using a proven analytical technique. Each dilution should be acidified with an appropriate high-purity acid and diluted to calibrated volume using high-purity water. The stability of the working standard solution will depend on the final acid concentration; therefore, care should be exercised to ensure that the final acid concentration of the dilution closely approximates that of the SRM. To achieve the highest accuracy, the analyst should prepare daily working solutions from 100 µg/mL dilutions of the original SRM solution.

REFERENCE

- [1] Taylor, B.N. and Kuyatt, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results" NIST Technical Note 1297, National Institute of Standards and Technology, Technology Administration, U.S. Department of Commerce, 1993.